

Manufacturer Declaration

- Safety in flight -

Löwenstein Medical Technology GmbH + Co. KG
Kronsaalsweg 40 ▪ 22525 Hamburg ▪ Germany

We hereby confirm that the following ventilation devices are therapeutic medical equipment:

prisma VENT30 (WM110TD)
prisma VENT30-C (WM110TD)
prisma VENT40 (WM110TD)
prisma VENT50 (WM120TD)
prisma VENT50-C (WM120TD)
prismaAQUA (WM100TH)

The devices are suitable for transport and use in aircrafts and comply with the requirements in accordance with **RTCA/DO-160G Section 21, Category M.**

Hamburg, June 2019

Thomas Weber
CQO, Director Quality Management and
Regulatory Affairs

